




## Separation of 18 PAH Compounds using HALO® PAH

231-P



### PEAK IDENTITIES

1. Naphthalene
2. Acenaphthylene
3. 1-methylnaphthalene
4. 2-methylnaphthalene
5. Acenaphthene
6. Fluorene
7. Phenanthrene
8. Anthracene
9. Fluoranthene
10. Pyrene
11. Benzo(a)anthracene
12. Chrysene
13. Benzo[b]fluoranthene
14. Benzo[k]fluoranthene
15. Benzo[a]pyrene
16. Dibenzo[a,h]anthracene
17. Benzo[g,h,i]perylene
18. Indeno[1,2,3-cd]pyrene

### TEST CONDITIONS:

**Column:** HALO 90 Å PAH, 2.7 µm, 3.0 x 100 mm

**Part Number:** 92843-612

**Mobile Phase A:** Water

**B:** Acetonitrile

Gradient:	Time	%B
	0.0	50
	8.0	100
	10.0	100

**Flow Rate:** 0.77 mL/min

**Initial Back Pressure:** 263 bar

**Temperature:** 30 °C

**Detection:** 280 nm

**Injection Volume:** 2 µL

**Sample Solvent:** Methanol

**Data Rate:** 100 Hz

**Response Time:** 0.025 sec

**Flow Cell:** 1 µL

**LC System:** Shimadzu Nexera X2

Polycyclic Aromatic Hydrocarbons (PAHs) are a group of more than 100 chemicals released from the combustion of coal, oil, gasoline, tobacco, and wood. They can also be found in cooked food. PAHs are persistent chemicals and must be closely regulated for early detection/monitoring to minimize hazardous exposure in the environment and/or use of contaminated raw materials in different industries. A rapid separation of the 16 compounds specified in EPA 610 and an additional 2 PAH compounds that are regularly analyzed is demonstrated on the HALO® PAH column showing excellent speed and resolution.

